



C-Star Calibration

Date	April 30, 2008	S/N#	CST-1118DR	Pathlength	25 cm
------	----------------	------	------------	------------	-------

Analog meter

V_d	0.061 V
V_{air}	4.826 V
V_{ref}	4.728 V

Temperature of calibration water	20.5 °C
Ambient temperature during calibration	25.4 °C

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x): $Tr = e^{-cx}$

To determine beam transmittance: $Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})$

To determine beam attenuation coefficient: $c = -1/x * \ln(Tr)$

V_d Meter output with the beam blocked. This is the offset.

V_{air} Meter output in air with a clear beam path.

V_{ref} Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V_{ref} .

Ambient temperature: meter temperature in air during the calibration.

V_{sig} Measured signal output of meter.